


Elements to Consider When Qualifying Control Environmental Chambers Used for CGMP Applications

May 21, 2013 2:41 pm EDT



As part of the installation, operational, and performance qualification (IQ, OQ, PQ) protocol of the control of environmental chambers, the thermo-mapping elements for testing are indicated as follows: Installation meeting the design specification
Empty chamber thermo-mapping test
Open-door study (for informational purposes only)
Power failure (for informational purposes only)
Full loaded chamber thermo-profile mapping. This is a technical article on how to perform a specific validation function. This article aims to give guidance to the industry on utilizing one methodology for thermo-mapping...

Source URL: <http://www.ivtnetwork.com/article/elements-consider-when-qualifying-control-environmental-chambers-used-cgmp-applications>